

## ☒ 7. Oxytetracycline

Column: PLRP-S 100Å 8μm, 250 × 4.6mm (1512-5800)  
Eluent: 94% 20mM K<sub>2</sub>HPO<sub>4</sub>, 1.5mM Bu<sub>4</sub>NHSO<sub>4</sub>,  
0.01mM EDTA (disodium salt), pH 9.0 : 6% *t*-butanol  
Flow Rate: 1.0ml/min  
Temp: 60°C  
Detector: UV, 210nm

