

☒31. Carbonated Soft Drink Analysis

KEY

- 1 Caffeine
- 2 Aspartame
- 3 Benzoic acid

Column: PLRP-S 100Å 5µm,
250 × 4.6mm (1512-5500)
Eluent : 70% 15mM Phosphoric
acid : 30% ACN
Flow Rate: 1.0ml/min
Detector: UV, 210nm

