

☒36. Separation of Standard Peptide Mixture

Column: PLRP-S 100Å 3µm, 150 × 4.6mm (1512-3300)
Eluent A: 0.1% TFA in 20% ACN
Eluent B: 0.1% TFA in 50% ACN
Gradient: 0 - 100% B in 15 mins
Flow Rate: 1.0ml/min
Detector: UV, 220nm

KEY

1. Oxytocin
2. Angiotensin II
3. Angiotensin I
4. Insulin

