

☒47. Polyethylene Glycols

Column: PLRP-S 100Å 5μm,
150 × 4.6mm (1111-3500)

Eluent A: Water

Eluent B: ACN

Gradient: 10 - 30% B in 12 mins
held at 30% B for 3 mins

Flow Rate: 1.0ml/min

Sample Conc: 1mg/ml

Detector: PL-ELS 2100(neb=50°C,
evap=70°C, gas=1.6 SLM)

