

☒64. Purity Assessment of an Affinity Separation

Column: PL-SAX 4000Å 8µm, 50 × 4.6mm (1551-1803)
Eluent A: 10mM Tris HCl, pH 8.0
Eluent B: A + 0.5M NaCl, pH 8.0
Gradient: 0 - 100% B in 2 mins
Flow Rate: 4.0ml/min
Detector: UV, 280nm

Peaks:
1. IgG
2. Transferrin
3. Albumin

